



EA MLA Signatory
Český institut pro akreditaci, o.p.s.
Hájkova 2747/22, Žižkov, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

CERTIFICATE OF ACCREDITATION

No. 113/2025

BMT Medical Technology s.r.o.
with registered office Cejl 157/50, Zábřdovice, 602 00 Brno,
Company Registration No. 46346996

for the Testing Laboratory No. 1325
Testing Laboratory of Steam and Hot Air Sterilizers and Heat Technology Devices

Scope of accreditation:

Tests of steam and hot air sterilizers and temperature control technology to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Conformity Assessment Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Accredited Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 589/2022 of 06/12/2022, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **10/03/2030**

Prague: 10/03/2025



Jan Velišek
Director of the Department
of Testing and Calibration Laboratories
Czech Accreditation Institute



**The Appendix is an integral part of
Certificate of Accreditation No. 113/2025 of 10/03/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

BMT Medical Technology s.r.o.

CAB number 1325, Testing Laboratory of Steam and Hot Air Sterilizers and Heat
Technology Devices
Cejl 157/50, Zábřehovice, 602 00 Brno

Tests:

Ordinal number ¹	Test procedure / method name	Test procedure / method identification ²	Tested subject	Degrees of freedom ³
1*	Air leakage test	QPP 10-52 (ČSN EN 285+A1, cl. 18; ČSN EN 13060+A1, cl. 10.2)	Steam sterilizers	-
2*	Thermometric test	QPP 10-53 (ČSN EN 285+A1, cl. 16)	Steam sterilizers	-
3*	Dynamic chamber pressure test	QPP 10-54 (ČSN EN 285+A1, cl. 22; ČSN EN 13060+A1, cl. 10.3)	Steam sterilizers	-
4*	Measurement of pressure and temperature profile in empty chamber	QPP 10-56 (ČSN EN 554:1996, cl. 5.2; VYR-34:2005, cl. 2.1)	Steam sterilizers	-
5*	Measurement of pressure and temperature distribution in chamber with the load	QPP 10-57 (ČSN EN 554:1996, cl. 5.3; VYR-34:2005, cl. 2.1)	Steam sterilizers	-
6*	Measurement of temperature	QPP 10-27 (ČSN EN IEC 60068-3-5 ed. 2; DIN 12 880)	Hot air sterilizers, heat technology devices	-
7*	Measurement of temperature profile in empty chamber	QPP 10-91 (VYR-34:2005, cl. 2.2)	Hot air sterilizers, heat technology devices	-
8*	Measurement of temperature distribution in chamber with the load	QPP 10-98 (VYR-34:2005, cl. 2.2)	Hot air sterilizers, heat technology devices	-
9*	Measurement of temperature and humidity	QPP 10-50 (ČSN EN IEC 60068-3-6 ed. 2; DIN 12 880)	Heat technology devices (climatic chambers)	-

- ¹ asterisk at the ordinal number identifies the tests, which the laboratory is qualified to carry out outside the permanent laboratory premises
- ² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)
- ³ the laboratory does not apply a flexible approach to the scope of accreditation

Explanations:

QPP - BMT Medical Technology s.r.o. Working Procedure

VYR - SÚKL Guideline (State Institute for Drug Control)

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself."

